Certificate of antimicrobial properties

CERTIFICATE

on the antimicrobial properties of

Nadicare Solutions GmbH TA 2219

of the company

Allcera B.V.
Westvoorstraat 8
3262 JP
Oud Beijerland
The Netherlands

has been successfully tested using QualiScreen SOP 3.5 (**S**tandard **O**peration **P**rocedure) "Test to determine the antimicrobial efficacy of materials against Staphylococcus epidermidis"

and is considered a n t i m i c r o b i a l

against Staphylococcus epidermidis DSM 18857

>99,9 %

For details, see Test Report No.: 1536 Rev 1 / 20130318-R07-01-51
Samples: 7501703130017

The bearer of this certificate is entitled to affix the following seal to parts or surfaces made from the tested material



Nuremberg, Germany, July 30th 2020

Managing Director

Laboratory Director

QualityLabs BT GmbH, Neumeyerstraße 46a, 90411 Nuremberg, Germany